



防爆合格证

CONFORMITY CERTIFICATE OF EXPLOSION-PROOF

证号
Certificate No. CE081562

产品名称
Name of Product
型号及规格
Type of Product
防爆标志
Marking
技术文件
Technical Documents
图号
Drawing No.
备注
Note (s)

气体检测探头

FGD10

Exd II CT4/T5/T6

/

SSC18/001 SSC18/002

1. 额定电压: DC24V;
2. 输出电流: 4~20mA;
3. 温度组别对应的环境温度:
T4: $-20^{\circ}\text{C} \leq T_a \leq 60^{\circ}\text{C}$ 、T5: $-20^{\circ}\text{C} \leq T_a \leq 50^{\circ}\text{C}$ 、
T6: $-20^{\circ}\text{C} \leq T_a \leq 40^{\circ}\text{C}$;
4. 制造商地址: Hermitage Lane Industrial Estate,
Kings Mill Way, Mansfield, Nottinghamshire,
England NG18 5ER, UK.

经对上述产品图样及技术文件的审查和样品的检验,其符合以下中国现行标准:

By verifying the drawings and technical documents and checking samples, the product complies with the following standards currently valid in P.R.China:

GB3836.1-2000

GB3836.2-2000

发给:
Issued to:

STATUS SCIENTIFIC CONTROLS LTD

本证失效日期:
Date of Expire:

2013-11-24

发证日期:
Date of Issue:

2008-11-24

中心印章
Center seal



中心主任
Director

本中心由以下组织认可:
PCEC has been approved by

石油和化学工业电气产品防爆质量监督检验中心
Supervision & Test Center of Ex-products of China Petroleum & Chemical Industry



注:本证仅对与送检样品一致的产品有效。

Note: This certificate is only valid for the products that are in accord with sample(s) tested and verified.

中心地址:中国天津市丁字沽三号路85号

邮政编码:300131

电话/传真:022-26651066/26689116

Center Add: No.85 No.3 Road DingZiGu Tianjin China Post code: 300131

Tel/ Fax: 022-26651066/26689116

E-mail:cec@pcec.com.cn

<http://www.pcec.com.cn>



中华人民共和国

计量器具型式批准证书

PATTERN APPROVAL CERTIFICATE OF THE MEASURING INSTRUMENTS OF THE PEOPLE'S REPUBLIC OF CHINA

英国 STATUS SCIENTIFIC CONTROLS LTD

根据《中华人民共和国计量法》及相关规定和技术要求，下列计量器具经定型鉴定合格，现予批准。

According to the Law on Metrology of the People's Republic of China and the relevant regulations, the pattern of measuring instruments applied for pattern approval have been approved.

计量器具名称及型号：

Name and type of the measuring instruments:

气体检测系统（探头 FGD10A/B；控制器 LC2/4/8 型）

计量器具的技术指标见型式注册表。

The technical specifications of the measuring instruments are described in the pattern registration list.

型式批准的标志与编号：

The mark and identification numbers of the pattern approval:



2009-C143

批准人

Approval signature

蔺长城

批准部门

Approval authority

批准日期 二〇〇九年四月六日

Approval date

批准时的附件：

1. 计量器具型式评价报告
2. 型式注册表
3. 型式评价系统

